



# Inspection Certificate

## PORTABLE HARDNESS TESTER

CODE: HDT-LP200

Standard Hardness Block (HLD)	Measured Value (HLD)						Error (HLD)
	1	2	3	4	5	Average Value	
523	521	521	522	523	524	522.2	-0.8
764	761	763	763	764	765	763.2	-0.8
802	799	801	801	801	803	801.0	-1.0

Serial No. E21084242000

Calibration temperature: 20°C

Humidity: 60%

Inspector:

*Case*