INSIZE

Inspection Certificate

SINGLE GAS DETECTOR **CODE: 0611-SG01**

Gas: CH_4 **Range: (0-100%) LEL**

Serial No. of main unit: 20260140843 **Release date: 01.2026**

Inspection Item	Technical requirements	Result
1.Appearance	Appearance should be smooth, neat	Qualified
2.Deviation	±5%FS	Qualified
3.Response time	Diffusion≤30s	Qualified
4.Alarm function	Sound, light, vibration alarm normal	Qualified
5.Alarm value	EX (%LEL)	Qualified
	20/50	
6.Calibration	EX (%LEL)	Qualified
	10/40/60	

Calibration temperature: 20°C Humidity: 60%RH

Inspector:

AEIMA