INSIZE

Inspection Certificate

MULTI GAS DETECTOR **CODE: 0612-MG04**

Serial No. of main unit: 20260240910 **Release date: 02.2026**

Inspection iter	Technical requirements				Result
Gas	CH_4	O_2	H_2S	CO	
1.Range	(0-100) %LEL	(0-30.0)%vol	(0−100) ppm	(0−1000) ppm	
2.Appearance	Appearance should be smooth, neat				Qualified
3.Alarm functio	Sound, light, vibration alarm normal				Qualified
4.Deviation	±5%FS	±0.7%VOL	±2ppm or ±10%	±5ppm or ±10%	Qualified
5.Response tim	⩽30s	⩽20s	⩽60s	⩽60s	Qualified
6.Alarm value	EX (%LEL)	O_2 (%VOL)	H_2S(ppm)	CO(ppm)	Qualified
	20/50	19.5/23.5	10/20	50/200	

Calibration temperature: 20°C Humidity: 60%RH

Inspector:

ДЕЛТА