

LARGE RANGE DIAL TEST INDICATOR

JEWELLED
BEARING

**INSPECTION
CERTIFICATE**
TRACEABLE TO NIST



2386-16A

- Dial reading: 0-40-0
- Jeweled bearing
- Carbide contact point
- Two measuring directions
- Anti magnetic body
- Supplied with two clamps:
diameter Ø4mm and Ø8mm
- Optional accessory: styli (page 182)

Code	Range	Graduation	Accuracy	Hysteresis
2386-16A*	1.6mm	0.01mm	25µm	8µm

*Supplied with manufacturer inspection certificate traceable to NIST USA

clamp (included)

