



KeySense Testing & Certification International Co., Ltd.
1-3F, Lab Building, No.29 District, Zhongkai Hi-Tech
Industrial Development Park, Huizhou, Guangdong,
China

Tel:0752-3219929
Fax:0752-3219929
<http://www.kst-cert.com>

Technical Compliance Statement

Certificate No.: KST231222E002Q01

Report No.: KST231222E002Q01

The following product has been tested by us with the listed standards and found in Compliance with the council EMC directive 2014/30/EU. It is demonstrative for the Compliance with this EMC Directive.

Holder: INSIZE CO., LTD.

Address: 80 Xiangyang Road, Suzhou New District, 215009 China

Product: FORCE GAGES

Model Name: ISF-DF5KA, ISF-DF, ISF-1DF, ISF-F, ISF-5M, ISF-5AM,
ISF-3M, ISF-QF, ISF-IF, ISF-KF, ISF-HF, ISF-RF, ISF-WF
(The appearance and range and reading display are different)

Trade Name: INSIZE

Test Standard	
EN 55011:2016+A2:2021	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement
EN IEC 61000-3-2:2019+A1: 2021	Electromagnetic Compatibility (EMC) Part 3-2:Limits-Limits for harmonic current emissions(equipment input current $\leq 16A$ per phase)
EN 61000-3-3:2013+A2:2 021	Electromagnetic Compatibility(EMC) Part 3-3:Limits-limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current $\leq 16A$ per phase and not subject to conditional connection
EN IEC 61326-1:2021	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements



Signature:

Name:


Position:

Date:

Jack.Li

Supervisor

Jan 04, 2024



The statement is based on single evaluation of one sample of above mentioned product. It does not imply an assessment of the whole production and does not permit the use of the test lab. Logo.