KeySense Testing & Certification International Co., Ltd.
1-3F,Lab Building, No.29 District, Zhongkai Hi-Tech Industrial Development Park, Huizhou, Guangdong, China

Tel:0752-3219929
Fax:0752-3219929
http://www.kst-cert.com

Technical Compliance Statement

Certificate No.: KST240201E001Q09
Report No.: KST240201E001Q09

The following product has been tested by us with the listed standards and found in Compliance with the council EMC directive 2014/30/EU. It is demonstrative for the Compliance with this EMC Directive.

Holder: INSIZE CO., LTD.
Address: 80 Xiangyang Road, Suzhou New District, 215009 China
Product: Scales/Balances
Model Name: 8304-120,8001,8002,8003,8004,8005,8006,8007,8008,8009,8010,8011,8012, 8013,8014,8015,8016,8017,8018,8019,8020,8101,8105,8107, 8108,8109,8110,8111,8112,8113,8114,8115,8116,8117,8118,8119,8120, 8301,8303,8304,8305,8306,8307,8308,8309,8310,8311,8312,8313, 8314,8315,8316,8317,8318,8319,8320,8401,8402,8403,8404,8405, 8406,8407,8408,8409,8410,8501,8502,8503,8504,8505,8506,8507, 8508,8509,8510,8601,8602,8603,8604,8605,8606,8607,8608,8609, 8610,8801,8802,8803,8804,8805,8806,8807,8808,8809,8810Series (Appearance, range, accuracy is different)
Trade Name: INSIZE

Test Standard	
EN 55011:2016+A2:2021	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement
EN IEC 61000-3-2:2019+A1:2021	Electromagnetic Compatibility (EMC) Part 3-2:Limits-Limits for harmonic current emissions(equipment input current ≦16A per phase)
EN 61000-3-3:2013+A2:2021	Electromagnetic Compatibility(EMC) Part 3-3:Limits-limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≦16A per phase and not subject to conditional connection
EN IEC 61326-1:2021	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

CE

Signature:
Name: Jack.Li
Position: Supervisor
Date: March 12, 2024

KeySense Testing & Certification International Co., Ltd.

The statement is based on single evaluation of one sample of above mentioned product. It does not imply an assessment of the whole production and does not permit the use of the test lab. Logo.