



# Inspection Certificate

## TORQUE WRENCH

CODE:IST-9WM335

Set Point (N.m)	Readings(N.m)			Error
65	65.16	64.89	64.94	0.2%
200	199.12	203.44	201.42	1.7%
335	338.33	335.58	336.77	1.0%

Serial No. AT139266

Calibration temperature: 20°C

Humidity: 60%

Inspector:

*Case*