

←INSIZE→

Inspection Certificate

ΔEITMA

Product name

Thread Ring Gage

Serial No.

0409250129

Code No.

4120-30

Standard temperature

20 °C

Size

M30x3.5 6g

Based on

ISO 1502 (1996)

Unit: mm

GO		Pitch Diameter	Minor Diameter
	Specified	27.662±0.012	26.158±0.012
	Observed	27.664	26.149

Judgement:

Passed

Inspector:

S. Yeng