

TESTING REPORT



Check:	Surface Flatness	Instrument:	Electronic Level
Size:	400x250x60mm	Graduation:	0.001 mm/m
Serial No. :	29040572		
Measured Data (Unit: μm)			
	<u>1.03</u>	0.75	0.34
	0.20	<u>0.00</u>	0.20
	0.34	0.69	<u>1.03</u>
Frame Drawing (Unit: μm)			
H=	1.03		
L=	0.00		
DIN 876 Grade:	00	0	1
Error allowed (mil) :	2.8	5.6	14.0
Result: Flatness=1.03 μm , meets DIN 876 Grade 00			
Inspector:	<i>S. Yeng</i>		