



Inspection Certificate

Unit: μm

Product name Dial Indicator
Code No. 2310-30A
Measuring range 30mm
Graduation 0.01mm
Serial No. 2916120
Standard temperature 20°C
Traceable to NIST No.821/279484-10

Maximum Permissible Error		Error
Any 1mm	15	6
Whole range	35	9
Hysteresis	7	6

Judgement: Passed

Inspector: *Opriul*

