



Inspection Certificate

Product name Digital Caliper

Code No. 1108-150

Measuring range 0-150mm / 0-6"

Resolution 0.01mm / 0.0005"

Serial No. 0206161840

Standard temperature 20 °C / 68 °F

Traceable to USA NIST No.821/279484-10

Germany DKD No.D-K-15190-01-00

Inspector

Judgement Passed

Unit: inch

Measuring Length	Permissible Error	Error	
		External	Internal
0.4 DIA	+0.0005 -0.0010	/	-0.0005
2	±0.0010	0.0005	/
4		0.0010	/
6	±0.0015	0.0005	/
8		/	/
10		/	/
12		/	/

Unit: mm

Measuring Length	Permissible Error	Error	
		External	Internal
Φ10	+0.01 -0.02	/	-0.01
50	±0.02	0.01	/
100		0.02	/
150	±0.03	0.01	/
200		/	/
250		/	/
300		/	/